

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II FS DNA Module
 Catalog Number: E7810L
 Packaging Lot Number: 10127290
 Expiration Date: 11/2022
 Storage Temperature: -20°C
 Specification Version: PS-E7810S/L v1.0

| NEBNext® Ultra™ II FS DNA Module Component List | | | |
|---|---------------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| E7808AAVIAL | TE Buffer | 10127293 | Pass |
| E7807AAVIAL | NEBNext® Ultra™ II FS Reaction Buffer | 10127292 | Pass |
| E7806AAVIAL | NEBNext® Ultra™ II FS Enzyme Mix | 10127291 | Pass |

| Assay Name/Specification | Lot # 10127290 |
|---|----------------|
| <p>Functional Testing (Library Construction, FS DNA) Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. A fragmentation time of 20 minutes was used to generate an insert size of approximately 200 bp. The final average library size is between 270 and 450 bp as determined by an Agilent Bioanalyzer. Libraries made from the previous and current lots for both input DNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p> | Pass |
| <p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II FS DNA Module and meet the designated specifications.</p> | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christine Sumner

Christine Sumner
Production Scientist
14 Dec 2021

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
28 Feb 2022