

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Amylose Resin

Catalog Number: E8021L
Packaging Lot Number: 10181436
Expiration Date: 01/2026
Storage Temperature: 4°C

Specification Version: PS-E8021S/L v2.0

Amylose Resin Component List			
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result
E8021LVIAL	Amylose Resin	10175671	Pass

Assay Name/Specification	Lot # 10181436
Functional Binding Assay (Resin Binding Capacity)	Pass
Amylose Resin (1 ml) was packed into a column and equilibrated with column buffer.	
Crude extract from E. coli containing a plasmid that expresses a	
MBP5*-paramyosinΔSal fusion protein ( 8 ml ) was then passed through the column at	
25°C, then washed with column buffer and the target protein eluted with ≥4 ml of	
column buffer containing 10 mM maltose. Binding capacity was determined to be >4 mg	
MBP5*-paramyosinΔSal /ml of resin based on A280 of the eluate.	
Functional Binding Assay (Resin Binding Specificity)	Pass
Amylose Resin (1 ml) was packed into a column and equilibrated with column buffer.	
Crude extract from E. coli containing a plasmid that expresses a	
MBP5*-paramyosinΔSal fusion protein (8 ml) was then passed through the column at	
25°C, and then washed with column buffer. The target protein was eluted with ≥4 ml	
of column buffer containing 10 mM maltose. SDS-PAGE of the eluate on a 10-20%	
Tris-Glycine gel confirms low non-specific binding of extract proteins and high	
isolation of target.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E8021L / Lot: 10181436 Page 1 of 2 Maxwell/Elkus Production Scientist 07 Feb 2023 Michael Tonello

Packaging Quality Control Inspector

09 Mar 2023

Page 2 of 2