

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Amylose Resin

Catalog Number: E8021S
Packaging Lot Number: 10102232
Expiration Date: 02/2024
Storage Temperature: 4°C

Specification Version: PS-E8021S/L v1.0

Amylose Resin Component List			
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result
E8021SVIAL	Amylose Resin	10097165	Pass

Assay Name/Specification	Lot # 10102232
Functional Binding Assay (Resin Binding Capacity)	Pass
Amylose Resin (1 ml) was packed into a column and equilibrated with column buffer.	
Crude extract from E. coli containing a plasmid that expresses a	
MBP5*-paramyosin $\Delta$ Sal fusion protein ( 8 ml ) was then passed through the column at	
25°C, then washed with column buffer and the target protein eluted with 4 ml of	
column buffer containing 10 mM maltose. Binding capacity was determined to be >4 mg	
MBP5*-paramyosinΔSal /ml of resin based on A280 of the eluate.	
Functional Binding Assay (Resin Binding Specificity)	Pass
Amylose Resin (1 ml) was packed into a column and equilibrated with column buffer.	
Crude extract from E. coli containing a plasmid that expresses a	
MBP5*-paramyosinΔSal fusion protein (8 ml) was then passed through the column at	
25°C, and then washed with column buffer. The target protein was eluted with 4 ml of	
column buffer containing 10 mM maltose. SDS-PAGE of the eluate on a 10-20%	
Tris-Glycine gel confirms low non-specific binding of extract proteins and high	
isolation of target.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E8021S / Lot: 10102232

Page 1 of 2

Michael Sproviero
Production Scientist

30 Mar 2021

Michael Tonello

Packaging Quality Control Inspector

30 Mar 2021

E8021S / Lot: 10102232

Page 2 of 2