

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	Amylose Resin High Flow
Catalog Number:	E8022S
Packaging Lot Number:	10230385
Expiration Date:	02/2027
Storage Temperature:	4°C
Specification Version:	PS-E8022S/L v2.0

Amylose Resin High Flow Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E8022SVIAL	Amylose Resin High Flow	10230384	Pass	

Assay Name/Specification	Lot # 10230385
Functional Binding Assay (Resin Binding Capacity) Amylose Resin High Flow (1 ml) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosin∆Sal fusion protein (8 ml) was then passed through the column at 25°C, then washed with column buffer and the target protein eluted with ≥4 ml of column buffer containing 10 mM maltose. Binding capacity was determined to be >4 mg MBP5*-paramyosin∆Sal /ml of resin based on A280 of the eluate.	Pass
Functional Binding Assay (Resin Binding Specificity) Amylose Resin High Flow (1 ml) was packed into a column and equilibrated with column buffer. Crude extract from E. coli containing a plasmid that expresses a MBP5*-paramyosin Δ Sal fusion protein (8 ml) was then passed through the column at 25°C, and then washed with column buffer. The target protein was eluted with ≥4 ml of column buffer containing 10 mM maltose. SDS-PAGE of the eluate on a 10-20% Tris-Glycine gel confirms low non-specific binding of extract proteins and high isolation of target.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Maxwell Elkus Production Scientist 23 Feb 2024

Michae on l

Michael Tonello Packaging Quality Control Inspector 06 Mar 2024

