

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTaq® DNA Polymerase

Catalog Number: M0480L Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 15

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Packaging Lot Number: 10082072
Expiration Date: 02/2022
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl , 100 mM KCl , 1 mM DTT , 0.1 mM EDTA , 0.5 % Tween®

20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0480S/L/X v2.0

OneTaq® DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0480LVIAL	OneTaq® DNA Polymerase	10067314	Pass	
B9026AVIAL	OneTaq® High GC Enhancer	10061963	Pass	
B9023SVIAL	OneTaq® GC Reaction Buffer	10071516	Pass	
B9022SVIAL	OneTaq® Standard Reaction Buffer	10061962	Pass	

Assay Name/Specification	Lot # 10082072
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of OneTaq® DNA Polymerase is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
PCR Amplification (Enhancer Dependent, >70% GC-rich) A 25 μl reaction in OneTaq® GC Reaction Buffer and 20% OneTaq® High GC Enhancer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® DNA Polymerase for 30 cycles of PCR amplification results in the enhancer-dependent production of the expected 627 bp product.	Pass
PCR Amplification (Buffer Dependent, >65% GC-rich) A 25 μl reaction in OneTaq® GC Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® DNA Polymerase for 30 cycles of PCR amplification results in the buffer-dependent production of the expected 737 bp product.	Pass



M0480L / Lot: 10082072

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 31 Aug 2020

hästie Vazguez

Josh Hersey

Packaging Quality Control Inspector

31 Aug 2020

