

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTaq® DNA Polymerase

Catalog Number: M0480S Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 15

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Packaging Lot Number: 10088137
Expiration Date: 04/2022
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 100 mM KCl, 1 mM DTT, 0.1 mM EDTA, 0.5 % Tween®

20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0480S/L/X v2.0

OneTaq® DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0480SVIAL	OneTaq® DNA Polymerase	10076363	Pass	
B9026AVIAL	OneTaq® High GC Enhancer	10061963	Pass	
B9023SVIAL	OneTaq® GC Reaction Buffer	10071516	Pass	
B9022SVIAL	OneTaq® Standard Reaction Buffer	10088290	Pass	

Assay Name/Specification	Lot # 10088137
PCR Amplification (5.0 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 5 ng Lambda DNA with 0.625 units of OneTaq® DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.	Pass
PCR Amplification (Buffer Dependent, >65% GC-rich) A 25 μl reaction in OneTaq® GC Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® DNA Polymerase for 30 cycles of PCR amplification results in the buffer-dependent production of the expected 737 bp product.	Pass
PCR Amplification (Enhancer Dependent, >70% GC-rich) A 25 μl reaction in OneTaq® GC Reaction Buffer and 20% OneTaq® High GC Enhancer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® DNA Polymerase for 30 cycles of PCR amplification results in the enhancer-dependent production of the expected 627 bp product.	Pass



M0480S / Lot: 10088137

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 30 Nov 2020

vistie Vazguez

Michael Tonello

Packaging Quality Control Inspector

30 Nov 2020

