

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTaq® DNA Polymerase

Catalog Number: M0480X
Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 15

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Packaging Lot Number: 10117473
Expiration Date: 01/2023
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 100 mM KCl, 1 mM DTT, 0.1 mM EDTA, 0.5 % Tween®

20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0480S/L/X v2.0

OneTaq® DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0480L	OneTaq® DNA Polymerase	10112815	Pass	

Assay Name/Specification	Lot # 10117473
PCR Amplification (5.0 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 5 ng Lambda DNA with 0.625 units of OneTaq® DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 2 containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA and a minimum of 5 units of OneTaq® DNA Polymerase incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of OneTaq® DNA Polymerase is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
PCR Amplification (Enhancer Dependent, >70% GC-rich) A 25 µl reaction in OneTaq® GC Reaction Buffer and 20% OneTaq® High GC Enhancer in	Pass



M0480X / Lot: 10117473

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 02 Aug 2021

production of the expected 737 bp product.

vistie Vanguez

Michael Tonello

Packaging Quality Control Inspector

02 Aug 2021



M0480X / Lot: 10117473

Page 2 of 2