

New England Biolabs Certificate of Analysis


Product Name: Phusion[®] High-Fidelity PCR Master Mix with HF Buffer
Catalog Number: M0531S
Concentration: 2 X Concentrate
Packaging Lot Number: 10206440
Expiration Date: 10/2024
Storage Temperature: -20°C
Specification Version: PS-M0531S/L v2.0
Composition (1X): 1 X Phusion[®] HF Buffer, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 20 units/ml Phusion[®] High-Fidelity DNA Polymerase

Phusion [®] High-Fidelity PCR Master Mix with HF Buffer Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0531SVIAL	Phusion [®] High-Fidelity PCR Master Mix with HF Buffer	10187514	Pass
B0515AVIAL	DMSO	10194246	Pass

Assay Name/Specification	Lot # 10206440
PCR Amplification (20 kb Lambda DNA, Master Mix) A 50 µl reaction in 1X Phusion [®] High-Fidelity PCR Master Mix with HF Buffer and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.	Pass
PCR Amplification (7.5 kb Human Genomic DNA, Master Mix) A 50 µl reaction in 1X Phusion [®] High-Fidelity PCR Master Mix with HF Buffer and 1.0 µM primers containing 50 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7.5 kb product.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.


Lea Antonopoulos
Production Scientist
07 Jun 2023


Josh Hersey
Packaging Quality Control Inspector
03 Oct 2023