

www.neb.com info@neb.com



New England Biolabs Certificate of Analysis

Product Name: NEBNext® High-Fidelity 2X PCR Master Mix

Catalog Number: M0541S

Concentration: 2 X Concentrate

Packaging Lot Number: 10266330
Expiration Date: 03/2026
Storage Temperature: -20°C

Specification Version: PS-M0541S/L v2.0

Composition (1X): Proprietary

| NEBNext® High-Fidelity 2X PCR Master Mix Component List | | | |
|---|--|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| M0541SVIAL | NEBNext® High-Fidelity 2X PCR Master Mix | 10266331 | Pass |

| Assay Name/Specification | Lot # 10266330 |
|---|----------------|
| Non-Specific DNase Activity (16 hour, Master Mix) A 50 µl reaction in 1X NEBNext® High-Fidelity 2X PCR Master Mix containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| PCR Amplification (Master Mix) A 50 μl reaction containing 0.5 μM primers with 20 ng human genomic DNA and 1X NEBNext® High-Fidelity 2X PCR Master Mix for 30 cycles of PCR amplification results in the expected 737 bp product. | Pass |
| Phosphatase Activity (pNPP) A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 20 µl of NEBNext® High-Fidelity 2X PCR Master Mix incubated for 4 hours at 37°C yields <0.00001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis. | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



M0541S / Lot: 10266330

Page 1 of 2



Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com



Christin Summ

Christine Sumner Production Scientist 19 Dec 2024 Michael Tonello

Packaging Quality Control Inspector

19 Dec 2024

M0541S / Lot: 10266330

Page 2 of 2