

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: WarmStart® Multi-Purpose LAMP/RT-LAMP 2X Master Mix (with UDG)

Catalog Number: M1708L

Concentration: 2 X Concentrate

Packaging Lot Number: 10146128
Expiration Date: 03/2024
Storage Temperature: -20°C

Specification Version: PS-M1708S/L v2.0

Composition (1X): Proprietary

WarmStart® Multi-Purpose LAMP/RT-LAMP 2X Master Mix (with UDG) Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M1708SVIAL	WarmStart® Multi-Purpose LAMP/RT-LAMP 2X Master Mix (with UDG)	10143686	Pass	

Assay Name/Specification	Lot # 10146128
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of WarmStart® Multi-Purpose LAMP/RT-LAMP 2X Master Mix (with UDG) is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of WarmStart® Multi-Purpose LAMP/RT-LAMP 2X Master Mix (with UDG) is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
Functional Testing (RT-LAMP, Master Mix) A 25 µl reaction with 1X WarmStart® Multi-Purpose LAMP/RT-LAMP Master Mix (with UDG) in the presence of 1X LAMP Primers containing 10 ng of genomic RNA and 1X LAMP fluorescent dye results in a threshold time of ≤ 20 minutes as determined by fluorescent detection.	Pass
Functional Testing (LAMP, Master Mix) A 25 µl reaction with 1X WarmStart® Multi-Purpose LAMP/RT-LAMP Master Mix (with UDG) in the presence of 1X LAMP Primers containing 10 ng genomic DNA and 1X LAMP	Pass



M1708L / Lot: 10146128

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 31 Mar 2022 Michael Tonello

Packaging Quality Control Inspector

31 Mar 2022

