

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: WarmStart® Colorimetric LAMP 2X Master Mix (DNA & RNA)

Catalog Number: M1800S

Concentration: 2 X Concentrate

Packaging Lot Number: 10055830 Expiration Date: 07/2020 Storage Temperature: -20°C

Specification Version: PS-M1800S/L v2.0

Composition (1X): Proprietary

| WarmStart® Colorimetric LAMP 2X Master Mix (DNA & RNA) Component List | | | | |
|---|--|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| M1800SVIAL | WarmStart® Colorimetric LAMP 2X Master Mix (DNA & RNA) | 10049944 | Pass | |

| Assay Name/Specification | Lot # 10055830 |
|---|----------------|
| Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X WarmStart® Colorimetric LAMP Master Mix (DNA & RNA) containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| Functional Testing (LAMP, Master Mix) A 25 µl reaction with 1X WarmStart® Colorimetric LAMP Master Mix (DNA & RNA) in the presence of 1X LAMP Primers containing 10 ng genomic DNA and 1X LAMP fluorescent dye results in a threshold time of ≤ 20 minutes as determined by fluorescent detection. | Pass |
| Functional Testing (RT-LAMP, Master Mix) A 25 μ l reaction with 1X WarmStart® Colorimetric LAMP Master Mix (DNA & RNA) in the presence of 1X LAMP Primers containing 10 ng of genomic RNA and 1X LAMP fluorescent dye results in a threshold time of \leq 20 minutes as determined by fluorescent detection. | Pass |
| qPCR DNA Contamination (E. coli Genomic) A minimum of 1 μl of WarmStart® Colorimetric LAMP Master Mix (DNA & RNA) is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome. | Pass |



M1800S / Lot: 10055830

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Christie Vazquez Production Scientist

29 Jul 2019

May Minichiello

Packaging Quality Control Inspector

02 Jan 2020



M1800S / Lot: 10055830

Page 2 of 2