

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Luna® Universal qPCR Master Mix

Catalog Number: M3003L

Concentration: 2 X Concentrate

Packaging Lot Number: 10093293
Expiration Date: 12/2022
Storage Temperature: -20°C

Specification Version: PS-M3003S/L/G/X/E v2.0

Composition (1X): Proprietary

| Luna® Universal qPCR Master Mix Component List | | | |
|--|---------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| M3003SVIAL | Luna® Universal qPCR Master Mix | 10090983 | Pass |

| Assay Name/Specification | Lot # 10093293 |
|--|----------------|
| Functional Testing (qPCR) Luna® Universal qPCR Master Mix is functionally tested in qPCR with human cDNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude. | Pass |
| Non-Specific DNase Activity (16 hour, Master Mix) A 50 µl reaction in 1X Luna® Universal qPCR Master Mix containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Luna® Universal qPCR Master Mix is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome. | Pass |
| RNase Activity Assay (4 Hour Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna® Universal qPCR Master Mix is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection. | Pass |



M3003L / Lot: 10093293

Page 1 of 2



This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 16 Dec 2020 Michael Tonello

Packaging Quality Control Inspector

16 Dec 2020