

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Luna® Universal qPCR Master Mix

Catalog Number: M3003X

Concentration: 2 X Concentrate

Lot Number: 10028235
Expiration Date: 09/2020
Storage Temperature: -20°C

Specification Version: PS-M3003S/L/G/X/E v1.0

Composition (1X): Proprietary

Luna® Universal qPCR Master Mix Component List			
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result
M3003L	Luna® Universal qPCR Master Mix	10025104	Pass

Assay Name/Specification	Lot # 10028235
Functional Testing (qPCR) Luna® Universal qPCR Master Mix is functionally tested in qPCR with human cDNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.	Pass
Non-Specific DNase Activity (16 hour, Master Mix) A 50 µl reaction in 1X Luna® Universal qPCR Master Mix containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Luna® Universal qPCR Master Mix is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
RNase Activity Assay (4 Hour Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna® Universal qPCR Master Mix is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass



M3003X / Lot: 10028235

Page 1 of 2

Josh Hersey

07 Nov 2018

Packaging Quality Control Inspector

This product has been tested and shown to be in compliance with all specifications.

Christie Vazquez Production Scientist

07 Nov 2018



M3003X / Lot: 10028235

Page 2 of 2