

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Luna® Universal Probe qPCR Master Mix

Catalog Number: M3004X

Concentration: 2 X Concentrate

Lot Number: 10045782
Expiration Date: 12/2020
Storage Temperature: -20°C

Specification Version: PS-M3004S/L/G/X/E v1.0

Composition (1X): Proprietary

Luna® Universal Probe qPCR Master Mix Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M3004L	Luna® Universal Probe qPCR Master Mix	10038445	Pass

Assay Name/Specification	Lot # 10045782
RNase Activity Assay (4 Hour Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna® Universal Probe qPCR Master Mix is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
Functional Testing (qPCR) Luna® Universal Probe qPCR Master Mix is functionally tested in qPCR with human cDNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.	Pass
Non-Specific DNase Activity (16 hour, Master Mix) A 50 µl reaction in 1X Luna® Universal Probe qPCR Master Mix containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 μl of Luna® Universal Probe qPCR Master Mix is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass



M3004X / Lot: 10045782 Page 1 of 2 Michael Tonello

23 May 2019

Packaging Quality Control Inspector



This product has been tested and shown to be in compliance with all specifications.

Christie Vazquez Production Scientist

23 May 2019

ISO 9001
Registered Quality Medical Devices

Medical Devices

M3004X / Lot: 10045782

Page 2 of 2