

New England Biolabs Certificate of Analysis

Product Name: *Thermostable USER[®] III Enzyme*

Catalog Number: *M5509S*

Concentration: *1,000 U/ml*

Unit Definition: *One unit is defined as the amount of enzyme required to nick 10 pmol of a 60-mer fluorescently labeled oligonucleotide duplex containing a single uracil base in 15 minutes at 65°C in a total reaction volume of 10 µL in 1X ThermoPol[®] Buffer.*

Packaging Lot Number: *10221936*

Expiration Date: *10/2025*

Storage Temperature: *-20°C*

Storage Conditions: *10 mM Tris-HCl, 50 mM KCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 178 µg/ml BSA (pH 7.4 @ 25°C)*


Specification Version: *PS-M5509S v1.0*

Thermostable USER [®] III Enzyme Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M5509SVIAL	Thermostable USER [®] III Enzyme	10211819	Pass
B9004SVIAL	ThermoPol [®] Reaction Buffer Pack	10201526	Pass

Assay Name/Specification	Lot # 10221936
<p>RNase Activity (Extended Digestion)</p> <p>A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Thermostable USER[®] III Enzyme is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.</p>	Pass
<p>qPCR DNA Contamination (E. coli Genomic)</p> <p>A minimum of 1 unit of Thermostable USER[®] III Enzyme is screened for the presence of E. coli genomic DNA using SYBR[®] Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Jamie Souza
Production Scientist
24 Oct 2023



Michael Tonello
Packaging Quality Control Inspector
08 Jan 2024