NEW ENGLAND BioLabs Inc.
be INSPIRED drive DISCOVERY stay GENUINE
240 County Road
Ipswich, MA 01938-2723
Tel 978-927-5054
Fax 978-921-1350
www.neb.com
info@neb.com

New England Biolabs
Certificate of Analysis

Product Name:	*Quick-Load® Taq 2X Master Mix*
Catalog Number:	*M0271L*
Concentration:	*2 X Concentrate*
Packaging Lot Number:	*10106744*
Expiration Date:	*12/2022*
Storage Temperature:	*-20°C*
Specification Version:	*PS-M0271S/L v2.0*
Composition (1X):	*10 mM Tris-HCl (pH 8.6 @ 25°C), 50 mM KCl, 1.5 mM MgCl2, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 5 % Glycerol, 0.08 % IGEPAL® CA-630, 0.05 % Tween® 20, 0.024 % Orange G, 0.0025 % Xylene cyanol, 33 units/ml Taq DNA Polymerase*

Quick-Load® Taq 2X Master Mix Component List			
NEB Part Number	**Component Description**	**Lot Number**	**Individual QC Result**
M0271SVIAL	Quick-Load® Taq 2X Master Mix	10092764	**Pass**
B9021SVIAL	Magnesium Chloride ($MgCl_2$) Solution	10092740	**Pass**

Assay Name/Specification	**Lot # 10106744**
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X Quick-Load® Taq Master Mix containing 1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37ºC results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	**Pass**
Single Stranded DNase Activity (FAM-Labeled Oligo) A 50 µl reaction in ThermoPol® Reaction Buffer containing a 10 nM solution of a fluorescent internal labeled oligonucleotide and a minimum of 25 units of Taq DNA Polymerase incubated for 30 minutes at 37°C and 75°C yields <10% degradation as determined by capillary electrophoresis.	**Pass**
Endonuclease Activity (Nicking) A 50 µl reaction in ThermoPol® Reaction Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 20 units of Taq DNA Polymerase incubated for 4 hours at 37°C and 75ºC results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	**Pass**
Phosphatase Activity (pNPP)	**Pass**

nqa ISO 9001 Registered Quality Management
nqa ISO 13485 Registered Medical Devices
nqa ISO 14001 Registered Environmental Management

NEW ENGLAND BioLabs Inc.
be INSPIRED drive DISCOVERY stay GENUINE
240 County Road
Ipswich, MA 01938-2723
Tel 978-927-5054
Fax 978-921-1350
www.neb.com
info@neb.com

Assay Name/Specification	Lot # 10106744
A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 100 units of Taq DNA Polymerase incubated for 4 hours at 37°C yields <0.0001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.	
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Quick-Load® Taq 2X Master Mix is incubated at 37ºC. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	**Pass**
Protein Purity Assay (SDS-PAGE) Taq DNA Polymerase is ≥ 99% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	**Pass**
qPCR DNA Contamination (E. coli Genomic) A minimum of 5 units of Taq DNA Polymerase is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	**Pass**
PCR Amplification (5 kb Lambda, Master Mix) A 25 µl reaction in 1X Quick-Load® Taq Master Mix and 0.2 µM primers containing 5 ng Lambda DNA for 25 cycles of PCR amplification results in the expected 5 kb product.	**Pass**

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez
Production Scientist
23 Apr 2021

Josh Hersey
Packaging Quality Control Inspector
23 Apr 2021

nqa ISO 9001 Registered Quality Management
nqa ISO 13485 Registered Medical Devices
nqa ISO 14001 Registered Environmental Management