

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® dsDNA Fragmentase®

Catalog Number: M0348L
Lot Number: 10055210
Expiration Date: 01/2021
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 50 mM NaCl, 0.1 mM EDTA, 50 % Glycerol, 0.15 %

Triton®X-100, 200 μg/ml BSA, (pH 7.5 @ 25°C)

Specification Version: PS-M0348S/L v1.0

| NEBNext® dsDNA Fragmentase® Component List |  |            |                      |  |
|--|--|------------|----------------------|--|
| <b>NEB Part Number</b>                     | Component Description                          | Lot Number | Individual QC Result |  |
| M0348AAVIAL                                | NEBNext® dsDNA Fragmentase®                    | 10049868   | Pass                 |  |
| B0511AAVIAL                                | MgCl2 Solution (200 mM)                        | 10049870   | Pass                 |  |
| B0349AAVIAL                                | NEBNext® dsDNA Fragmentase® Reaction Buffer v2 | 10049869   | Pass                 |  |

| Assay Name/Specification   | Lot # 10055210 |
|--|----------------|
| Protease Activity (SDS-PAGE) A 20 µl reaction in 1X NEBNext® dsDNA Fragmentase Reaction Buffer containing 24 µg of a standard mixture of proteins and a minimum of 10 µl of NEBNext® dsDNA Fragmentase® incubated for 20 hours at 37°C, results in no detectable degradation of the protein mixture as determined by SDS-PAGE with Coomassie Blue detection. | Pass           |
| Phosphatase Activity (pNPP) A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 10 µl of NEBNext® dsDNA Fragmentase® incubated for 4 hours at 37°C yields <0.00001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.                            | Pass           |
| Protein Purity Assay (SDS-PAGE)  NEBNext® dsDNA Fragmentase® is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.  | Pass           |

This product has been tested and shown to be in compliance with all specifications.



M0348L / Lot: 10055210

Page 1 of 2

Christin Summ

Christine Sumner Production Scientist 05 Sep 2019 Michael Tonello

Packaging Quality Control Inspector

05 Sep 2019