

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Quick-Load® 100 bp DNA Ladder
Catalog Number:	N0467L
Concentration:	50 μg/ml
Unit Definition:	N/A
Lot Number:	10010027
Expiration Date:	03/2020
Storage Temperature:	4°C
Storage Conditions:	3.3 mM Tris-HCI (pH 8.0), 11 mM EDTA , 0.015 % Bromophenol Blue ,
	0.017 % SDS , 2.5 % Ficoll 400
Specification Version:	PS-N0467S/L v1.0

Quick-Load® 100 bp DNA Ladder Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N0467SVIAL	Quick-Load® 100 bp DNA Ladder	0561803	Pass	

Assay Name/Specification	Lot # 10010027
<b>DNA Concentration (A260)</b> The concentration of Quick-Load <sup>®</sup> 100 bp DNA Ladder is between 50 and 55 $\mu$ g/ml as determined by UV absorption at 260 nm.	Pass
<b>Electrophoretic Pattern (Marker)</b> The banding pattern of Quick-Load® 100 bp DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	Pass
Non-Specific DNase Activity (DNA, 16 hour) A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of Quick-Load® 100 bp DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.





*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Ana Egana Production Scientist 23 May 2018

Michae 711.

Michael Tonello Packaging Quality Control Inspector 23 May 2018

