

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Quick-Load® 100 bp DNA Ladder

Catalog Number:N0467LConcentration:50 μg/mlUnit Definition:N/A

Packaging Lot Number: 10185487
Expiration Date: 04/2025
Storage Temperature: 4°C

Storage Conditions: 3.3 mM Tris-HCl (pH 8.0), 11 mM EDTA, 0.015 % Bromophenol Blue,

0.017 % SDS, 2.5 % Ficoll 400

Specification Version: PS-N0467S/L v1.0

Quick-Load® 100 bp DNA Ladder Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N0467SVIAL	Quick-Load® 100 bp DNA Ladder	10185485	Pass

Assay Name/Specification	Lot # 10185487
DNA Concentration (A260) The concentration of Quick-Load® 100 bp DNA Ladder is between 50 and 55 μg/ml as	Pass
determined by UV absorption at 260 nm.	
Electrophoretic Pattern (Marker)	Pass
The banding pattern of Quick-Load® 100 bp DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with	
Ethidium Bromide at a concentration of 0.5 μ g/ml.	
Non-Specific DNase Activity (DNA, 16 hour)	Pass
A 50 μl reaction in 1X NEBuffer 2 containing 2.5 μg of Quick-Load® 100 bp DNA Ladder	
incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease	
degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N0467L / Lot: 10185487

Page 1 of 2

Vanessa Mathieu-Sheltry **Production Scientist** 04 Apr 2023

Josh Hersey Packaging Quality Control Inspector

28 Jun 2023