

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Quick-Load® Purple 100 bp DNA Ladder

Catalog Number:N0551LConcentration:50 μg/mlUnit Definition:N/A

Packaging Lot Number: 10091983
Expiration Date: 04/2022
Storage Temperature: 4°C

Storage Conditions: 2.5 % Ficoll 400 , 10 mM EDTA , 3.3 mM Tris-HCl (pH 8.0), 0.001 %

Dye 2 , 0.02 % Dye 1

Specification Version: PS-N0551S v1.0

| Quick-Load® Purple 100 bp DNA Ladder Component List | | | | |
|---|--------------------------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| N0551SVIAL | Quick-Load® Purple 100 bp DNA Ladder | 10070827 | Pass | |
| B7025SVIAL | Gel Loading Dye, Purple (6X), no SDS | 10074977 | Pass | |

| Assay Name/Specification | Lot # 10091983 |
|--|----------------|
| DNA Concentration (A260) | Pass |
| The concentration of Quick-Load® Purple 100 bp DNA Ladder is between 50 and 55 µg/ml as determined by UV absorption at 260 nm. | |
| Electrophoretic Pattern (Marker) | Pass |
| The banding pattern of Quick-Load® Purple 100 bp DNA Ladder on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml. | |
| Non-Specific DNase Activity (DNA, 16 hour) | Pass |
| A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of Quick-Load® Purple 100 bp DNA | |
| Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable | |
| nuclease degradation as determined by agarose gel electrophoresis. | |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N0551L / Lot: 10091983

Page 1 of 2



Ana Egana
Production Scientist
21 Dec 2020

Michael Tonello

Packaging Quality Control Inspector

21 Dec 2020