

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Lambda DNA-Hind III Digest

Catalog Number:N3012LConcentration:500 μg/mlUnit Definition:N/A

Packaging Lot Number: 10110118
Expiration Date: 03/2023
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3012S/L v1.0

Lambda DNA-Hind III Digest Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N3012LVIAL	Lambda DNA-Hind III Digest	10101807	Pass
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10093119	Pass

Assay Name/Specification	Lot # 10110118
Electrophoretic Pattern (Marker) The banding pattern of λ DNA-HindIII Digest on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 μ g/ml.	Pass
DNA Concentration (A260) The concentration of λ DNA-HindIII Digest is between 500 and 550 μ g/ml as determined by UV absorption at 260 nm.	Pass
A260/A280 Assay The ratio of UV absorption of λ DNA-HindIII Digest at 260 and 280 nm is between 1.8 and 2.0.	Pass
Non-Specific DNase Activity (DNA, 16 hour) A 50 μl reaction in 1X NEBuffer 2 containing 2.5 μg of λ DNA-HindIII Digest incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N3012L / Lot: 10110118 Page 1 of 2



Ana Egana Production Scientist 14 Jun 2021 Josh Hersey

14 Jun 2021

Packaging Quality Control Inspector

nga. ISO 9001 SO 13485 Registered Quality Medical Devices Management