

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Lambda DNA-BstE II Digest

Catalog Number: N3014S
Concentration: 500 μg/ml

Unit Definition: N/A

Packaging Lot Number: 10098962
Expiration Date: 02/2023
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3014S/L v1.0

Lambda DNA-BstE II Digest Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
N3014SVIAL	Lambda DNA-BstE II Digest	10098963	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10093119	Pass	

Assay Name/Specification	Lot # 10098962
Non-Specific DNase Activity (DNA, 16 hour) A 50 μl reaction in 1X NEBuffer 2 containing 2.5 μg of λ DNA-BstEII Digest incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
A260/A280 Assay The ratio of UV absorption of λ DNA-BstEII Digest at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) The concentration of λ DNA-BstEII Digest is between 500 and 550 μg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Marker) The banding pattern of $\lambda$ DNA-BstEII Digest on a 1.2% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 $\mu$ g/ml.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N3014S / Lot: 10098962



Ana Egana **Production Scientist** 12 Feb 2021

Josh Hersey Packaging Quality Control Inspector

12 Feb 2021