

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Lambda DNA-Mono Cut Mix

Catalog Number: N3019S Concentration: 500 μg/ml

Unit Definition: N/A

Packaging Lot Number: 10144809
Expiration Date: 03/2024
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3019S/L v1.0

Lambda DNA-Mono Cut Mix Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
N3019SVIAL	Lambda DNA-Mono Cut Mix	10144808	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10123372	Pass	

Assay Name/Specification	Lot # 10144809
Non-Specific DNase Activity (DNA, 16 hour) A 50 μl reaction in 1X NEBuffer 2 containing 2.5 μg of λ DNA-Mono Cut Mix incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
A260/A280 Assay The ratio of UV absorption of λ DNA-Mono Cut Mix at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) The concentration of $\lambda$ DNA-Mono Cut Mix is between 500 and 550 $\mu$ g/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (PFGE Marker) The banding pattern of λ DNA-Mono Cut Mix on a 1% CHEF PFG gel shows discrete, clearly identifiable bands at each band of the marker, as determined by gel electrophoresis using Ethidium Bromide.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N3019S / Lot: 10144809 Page 1 of 2



Vanessa Mathieu-Sheltry Production Scientist 20 Apr 2022

Erin Varney

Packaging Quality Control Inspector

20 Apr 2022

N3019S / Lot: 10144809

Page 2 of 2