

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: 1 kb Plus DNA Ladder

Catalog Number:N3200LConcentration:1,000 μg/mlPackaging Lot Number:10197910Expiration Date:03/2025Storage Temperature:-20°C

Specification Version: PS-N3200S/L v2.0

1 kb Plus DNA Ladder Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N3200LVIAL	1 kb Plus DNA Ladder	10183742	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10186772	Pass	

Assay Name/Specification	Lot # 10197910
A260/A280 Assay	Pass
The ratio of UV absorption of 1 kb Plus DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	
DNA Concentration (A260)	Pass
The concentration of 1 kb Plus DNA Ladder is between 1000 and 1050 µg/ml as determined by UV absorption at 260 nm.	
Electrophoretic Pattern (Marker)	Pass
The banding pattern of 1 kb Plus DNA Ladder on a 1.2% agarose gel shows discrete,	
clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 µg/ml.	
Diomide at a concentration of 0.5 μg/mi.	
Non-Specific DNase Activity (DNA, 16 hour)	Pass
A 50 μl reaction in 1X NEBuffer 2 containing 5 μg of 1 kb Plus DNA Ladder incubated	
for 16 hours at 37°C results in a DNA pattern free of detectable nuclease	
degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N3200L / Lot: 10197910

Page 1 of 2

Vanessa Mathieu-Sheltry Production Scientist

21 Mar 2023

Michael Tonello

Packaging Quality Control Inspector 07 Jul 2023

nqa.

ISO 9001
Registered
Quality
Management
Medical Device



