

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	HeLa Genomic DNA
Catalog Number:	N4006S
Concentration:	100 μg/ml
Unit Definition:	N/A
Lot Number:	10009459
Expiration Date:	05/2020
Storage Temperature:	-20°C
Storage Conditions:	10 mM Tris-HCl , 1 mM EDTA, (pH 7.5 @ 25°C)
Specification Version:	PS-N4006S v1.0

HeLa Genomic DNA Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N4006SVIAL	HeLa Genomic DNA	10009460	Pass	

Assay Name/Specification	Lot # 10009459
<b>A260/A280 Assay</b> The ratio of UV absorption of HeLa Genomic DNA at 260 and 280 nm is between 1.8 and 2.0.	Pass
<b>DNA Concentration (A260)</b> The concentration of HeLa Genomic DNA is between 100 and 110 $\mu$ g/ml as determined by UV absorption at 260 nm.	Pass
<b>Electrophoretic Pattern (Genomic DNA)</b> The banding pattern of HeLa Genomic DNA on a 1.2% agarose gel is evaluated against a control lot for relative integrity and intensity as determined by gel electrophoresis using Ethidium Bromide.	Pass
<b>Non-Specific DNase Activity (Genomic DNA, 16 hour)</b> A 50 µl reaction in 1X NEBuffer 2 containing 2.5 µg of HeLa Genomic DNA incubated for 16 hours at 37°C does not produce any further detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Restriction Digest (Genomic DNA)</b> A 50 µl reaction in NEBuffer 2.1 containing 2.5 µg of HeLa Genomic DNA and 20 units of HindIII incubated for 1 hour at 37°C produces the expected fragmentation pattern as determined by agarose gel electrophoresis.	Pass





*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

This product has been tested and shown to be in compliance with all specifications.

Nulhice

Vanessa Mathieu-Sheltry Production Scientist 23 May 2018

Mich 71.11

Michael Tonello Packaging Quality Control Inspector 23 May 2018

