

## New England Biolabs Certificate of Analysis


*Product Name:* pTYB21 Vector  
*Catalog Number:* N6709S  
*Concentration:* 200 µg/ml  
*Packaging Lot Number:* 10243986  
*Expiration Date:* 11/2026  
*Storage Temperature:* -20°C  
*Storage Conditions:* 10 mM Tris-HCl, 1 mM EDTA, (pH 8.0 @ 25°C)  
*Specification Version:* PS-N6709S v1.0

pTYB21 Vector Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N6709SVIAL	pTYB21 Vector	10217197	Pass

Assay Name/Specification	Lot # 10243986
<b>A260/A280 Assay</b> The ratio of UV absorption of pTYB21 Vector at 260 and 280 nm is between 1.8 and 2.0.	<b>Pass</b>
<b>DNA Concentration (A260)</b> The concentration of pTYB21 Vector is between 200 and 220 µg/ml as determined by UV absorption at 260 nm.	<b>Pass</b>
<b>Restriction Digest (Linearization)</b> A 50 µl reaction in CutSmart® Buffer containing 2 µg of pTYB21 Vector and 10 units of SapI incubated for 1 hour at 37°C produces > 95% linearization resulting in a single band of approximately 7514 bp as determined by agarose gel electrophoresis.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



---

Brad Landgraf  
Production Scientist  
21 Nov 2023



---

Josh Hersey  
Packaging Quality Control Inspector  
26 Jul 2024