

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Color Prestained Protein Standard, Broad Range (10-250 kDa)

Catalog Number: P7719S
Lot Number: 10040351
Expiration Date: 03/2021
Storage Temperature: -20°C

Storage Conditions: 62.5 mM Tris-Phosphate, 1 mM EDTA, 2% SDS, 10 mM DTT, 1 mM NaN3, 33%

Glycerol

Specification Version: PS-P7719S/L v1.0

| Color Prestained Protein Standard, Broad Range (10-250 kDa) Component List | | | | |
|--|---|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| P7719SVIAL | Color Prestained Protein Standard, Broad Range (10-250 kDa) | 10040349 | Pass | |

| Assay Name/Specification | Lot # 10040351 |
|--|----------------|
| Electrophoretic Pattern (Protein Marker) | Pass |
| The banding pattern of Color Prestained Protein Standard, Broad Range (10-250 kDa) | |
| on a 10-20% Tris-Glycine gel shows discrete, clearly identifiable bands at each band | |
| of the marker, as determined by gel electrophoresis using Coomassie Blue detection. | |
| Protease Activity (Protein Markers & Ladders) | Pass |
| | Fass |
| A 20 µl sample of Color Prestained Protein Standard, Broad Range (10-250 kDa) | |
| incubated for 16 hours at 37°C, results in no detectable degradation of the protein | |
| mixture as determined by SDS-PAGE with Coomassie Blue detection. | |

This product has been tested and shown to be in compliance with all specifications.



P7719S / Lot: 10040351 Page 1 of 2 Alicia Bielik **Production Scientist** 19 Mar 2019

Josh Hersey Packaging Quality Control Inspector

21 Mar 2019

P7719S / Lot: 10040351

Page 2 of 2