

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: EcoRI
Catalog Number: R0101L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10060022 Expiration Date: 11/2021 Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM KPO4 (pH 7.5), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 0.15% Triton X-100, 200 µg/ml BSA

Specification Version: PS-R0101S/L v2.0

EcoRI Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0101LVIAL	EcoRI	10060020	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10055732	Pass	
B0101SVIAL	NEBuffer™ EcoRI	0091905	Pass	

Assay Name/Specification	Lot # 10060022
Blue-White Screening (Terminal Integrity)	Pass
A sample of pUC19 vector linearized with a 10-fold excess of EcoRI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	
Protein Purity Assay (SDS-PAGE)	Pass
EcoRI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	
Non-Specific DNase Activity (16 Hour)	Pass
A 50 µl reaction in EcoRI Buffer containing 1 µg of Lambda DNA and a minimum of 100	
Units of EcoRI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	
Ligation and Recutting (Terminal Integrity)	Pass
After a 20-fold over-digestion of Lambda DNA with EcoRI, >95% of the DNA fragments	
can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with EcoRI.	



R0101L / Lot: 10060022

Page 1 of 2

Assay Name/Specification	Lot # 10060022
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in EcoRI Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 200 units of EcoRI incubated for	Pass
4 hours at 37°C releases <0.1% of the total radioactivity.	

This product has been tested and shown to be in compliance with all specifications.

Stephanie Cornelio Production Scientist

21 Nov 2019

Michael Tonello

Packaging Quality Control Inspector

18 Dec 2019

