

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Mspl
Catalog Number: R0106L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10093156
Expiration Date: 11/2022
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0106S/L v1.0

Mspl Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R0106LVIAL	Mspl	10089426	Pass	
B7204SVIAL	CutSmart® Buffer	10091033	Pass	
B7024AVIAL	Gel Loading Dye, Purple (6X)	10089401	Pass	

Assay Name/Specification	Lot # 10093156
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of Mspl incubated for 16 hours at 37°C results in a DNA pattern free of	Pass
detectable nuclease degradation as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 100 units of Mspl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity)  After a 20-fold over-digestion of Lambda DNA with MspI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with MspI.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



R0106L / Lot: 10093156 Page 1 of 2



Mala-Samaranayake Production Scientist 21 Dec 2020

Michael Tonello

Packaging Quality Control Inspector

21 Dec 2020