

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Kpnl
Catalog Number: R0142L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10058271
Expiration Date: 09/2021
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0142S/L v2.0

Kpnl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0142LVIAL	Kpnl	10055765	Pass	
B7201SVIAL	NEBuffer™ 1.1	10043905	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10046085	Pass	

Assay Name/Specification	Lot # 10058271
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 1.1 containing 1 µg of a mixture of single and	Pass
double-stranded [³H] E. coli DNA and a minimum of 100 units of KpnI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of pXba DNA with KpnI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with KpnI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 1.1 containing 1 µg of pXba DNA and a minimum of 50 Units of KpnI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) KpnI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass



R0142L / Lot: 10058271

Page 1 of 2

to the nicked form as determined by agarose gel electrophoresis.

This product has been tested and shown to be in compliance with all specifications.

Savid Guo

Production Scientist

25 Jul 2019

Michael Tonello

Packaging Quality Control Inspector

13 Nov 2019



R0142L / Lot: 10058271

Page 2 of 2