

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Xhol
Catalog Number: R0146L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA (HindIII digest) fragments in 1 hour at 37°C in a

total reaction volume of 50 µl.

Packaging Lot Number: 10107344
Expiration Date: 05/2023
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0146S/L v2.0

Xhol Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0146LVIAL	Xhol	10107342	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10093123	Pass
B6004SVIAL	rCutSmart™ Buffer	10107576	Pass

Assay Name/Specification	Lot # 10107344
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda HindIII DNA and a minimum of 100 Units of Xhol incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pXba DNA with XhoI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with XhoI.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of Xhol incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Protein Purity Assay (SDS-PAGE)	Pass



R0146L / Lot: 10107344

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Penghaa Zhang Production Scientist

21 May 2021

Michael Tonello

Packaging Quality Control Inspector

21 May 2021



R0146L / Lot: 10107344

Page 2 of 2