

New England Biolabs Certificate of Analysis


Product Name: XhoI
Catalog Number: R0146S
Concentration: 20,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (HindIII digest) fragments in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10035371
Expiration Date: 12/2020
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0146S/L v2.0

XhoI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0146SVIAL	XhoI	10032181	Pass
B7204SVIAL	CutSmart® Buffer	10021121	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10021140	Pass

Assay Name/Specification	Lot # 10035371
Blue-White Screening (Terminal Integrity) A sample of Litmus 28i vector linearized with a 10-fold excess of XhoI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled pBR322 DNA and a minimum of 100 Units of XhoI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of XhoI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pXba DNA with XhoI, >95% of the DNA fragments can	Pass

Assay Name/Specification	Lot # 10035371
<p>be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with XhoI.</p> <p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda HindIII DNA and a minimum of 100 Units of XhoI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p> <p>Protein Purity Assay (SDS-PAGE) XhoI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	<p style="text-align: center;">Pass</p> <p style="text-align: center;">Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Loren Duquette
Production Scientist
06 Feb 2019



Michael Tonello
Packaging Quality Control Inspector
11 Feb 2019