

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	Hinfl
Catalog Number:	R0155S
Concentration:	10,000 U/ml
Unit Definition:	One unit is defined as the amount of enzyme required to digest 1 μg of Lambda DNA in 1 hour at 37°C in a total reaction of 50 μl.
Lot Number:	10030643
Expiration Date:	11/2020
Storage Temperature:	-20°C
Storage Conditions:	50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 μg/ml BSA
Specification Version:	PS-R0155S/L v1.0

Hinfl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0155SVIAL	Hinfl	10030644	Pass	
B7204SVIAL	CutSmart® Buffer	10021125	Pass	

Assay Name/Specification	Lot # 10030643
Protein Purity Assay (SDS-PAGE) Hinfl is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of Hinfl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with Hinfl, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Hinfl.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Hinfl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass

This product has been tested and shown to be in compliance with all specifications.





be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

tephonie Onetto

Stephanie Cornelio Production Scientist 30 Nov 2018

Michae 11.

Michael Tonello Packaging Quality Control Inspector 02 Jan 2019

