

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Dpnl
Catalog Number: R0176L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pBR322 DNA (dam methylated) in rCutSmart Buffer in 1 hour at 37°C

in a total reaction volume of 50 μl.

Packaging Lot Number: 10198126
Expiration Date: 07/2025
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 400 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol,

200 μg/ml rAlbumin (pH 7.4 @ 25°C)

Specification Version: PS-R0176S/L v2.0

DpnI Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0176LVIAL	DpnI	10196895	Pass	
B7024AVIAL	Gel Loading Dye, Purple (6X)	10184699	Pass	
B6004SVIAL	rCutSmart™ Buffer	10196035	Pass	

Assay Name/Specification	Lot # 10198126
Endonuclease Activity (Nicking) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 20 units of DpnI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 200 units of DpnI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of pBR322 DNA with DpnI, ~25% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with DpnI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of pBR322 DNA and a minimum of	Pass



R0176L / Lot: 10198126

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

YunJie Sun

Production Scientist

11 Jul 2023

Michael Tonello

Packaging Quality Control Inspector

14 Jul 2023



R0176L / Lot: 10198126

Page 2 of 2