

## New England Biolabs Certificate of Analysis

**Product Name:** Xmal  
**Catalog Number:** R0180M  
**Concentration:** 50,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg adenovirus-2 in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10069147  
**Expiration Date:** 03/2022  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0180M v1.0

Xmal Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0180MVIAl	Xmal	10069148	Pass
B7204SVIAl	CutSmart® Buffer	10074633	Pass
B7024SVIAl	Gel Loading Dye, Purple (6X)	10071081	Pass

Assay Name/Specification	Lot # 10069147
<b>Protein Purity Assay (SDS-PAGE)</b> Xmal is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 10 Units of Xmal incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 10-fold over-digestion of pXba DNA with Xmal, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Xmal.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 50 units of Xmal incubated for 4 hours at 37°C releases <0.2% of the total radioactivity.	Pass

Assay Name/Specification	Lot # 10069147
<p><b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 50 Units of XmaI incubated for 4 hours at 37°C results in &lt;10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.




---

Penghua Zhang  
Production Scientist  
18 Jun 2020




---

Michael Tonello  
Packaging Quality Control Inspector  
18 Jun 2020