

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Ncil
Catalog Number: R0196L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10069837
Expiration Date: 03/2022
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0196S/L v1.0

Ncil Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0196LVIAL	Ncil	10069836	Pass	
B7204SVIAL	CutSmart® Buffer	10074633	Pass	

Assay Name/Specification	Lot # 10069837
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and	Pass
double-stranded [³H] E. coli DNA and a minimum of 200 units of Ncil incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Ncil, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Ncil.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 20 Units of Ncil incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) Ncil is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.



R0196L / Lot: 10069837 Page 1 of 2



Penghaa Zhang Production Scientist 23 Jun 2020 Michael Tonello

Michael Tonello

Packaging Quality Control Inspector 23 Jun 2020