

240 County Road Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: BstBl
Catalog Number: R0519L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 65°C in a total reaction volume of 50 μl.

Lot Number: 1001962: Expiration Date: 08/2020 Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 μg/ml BSA

Specification Version: PS-R0519S/L v1.0

BstBl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0519LVIAL	BstBl	10019623	Pass	
B7204SVIAL	CutSmart® Buffer	10015394	Pass	

Assay Name/Specification	Lot # 10019622
Endonuclease Activity (Nicking) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of supercoiled pUC19 DNA and a minimum of 20 units of BstBl incubated for 4 hours at 65°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BstBI incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with BstBI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 4 hours at 25°C. Of these ligated fragments, >95% can be recut with BstBI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 20 units of BstBl incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



R0519L / Lot: 10019622

Page 1 of 2



This product has been tested and shown to be in compliance with all specifications.

Tony Spear-Alfonso Production Scientist

23 Jul 2018

Michael Tonello

Packaging Quality Control Inspector

22 Aug 2018