

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Xcml
Catalog Number: R0533L
Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Lot Number: 1002705
Expiration Date: 10/2020
Storage Temperature: -20°C

Storage Conditions: 250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 0.15% Triton X-100, 200 µg/ml BSA

Specification Version: PS-R0533S/L v1.0

Xcml Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0533LVIAL	Xcml	10027053	Pass	
B7202SVIAL	NEBuffer™ 2.1	0261805	Pass	

Assay Name/Specification	Lot # 10027052
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in NEBuffer 2.1 containing 1 µg of a mixture of single and double-stranded [3H] E. coli DNA and a minimum of 50 units of Xcml incubated for 4	
hours at 37°C releases <0.1% of the total radioactivity.	
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Xcml, ~25% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Xcml.	Pass
Non-Specific DNase Activity (16 hour)	Pass
A 50 µl reaction in NEBuffer 2.1 containing 1 µg of Lambda DNA and a minimum of 5	
Units of Xcml incubated for 16 hours at 37°C results in a DNA pattern free of	
detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended	
incubations and/or high concentrations of this enzyme may result in star activity.	
See the product FAQ for recommended reaction conditions for this enzyme.	

This product has been tested and shown to be in compliance with all specifications.



R0533L / Lot: 10027052

Page 1 of 2



Jianying Luo **Production Scientist**

Michael Tonello

Packaging Quality Control Inspector 05 Nov 2018



01 Nov 2018



R0533L / Lot: 10027052

Page 2 of 2