

## New England Biolabs Certificate of Analysis

**Product Name:** *MscI*  
**Catalog Number:** *R0534L*  
**Concentration:** *5,000 U/ml*  
**Unit Definition:** *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in rCutSmart Buffer in 1 hour at 37°C in a total reaction volume of 50 µl.*  
**Packaging Lot Number:** *10282913*  
**Expiration Date:** *03/2027*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *10 mM Tris-HCl, 100 mM KCl, 1 mM DTT, 0.1 mM EDTA, 0.1 % Triton®X-100, 50 % Glycerol, 200 µg/mL rAlbumin (pH 7.4 @ 25°C)*  
**Specification Version:** *PS-R0534S/L v4.0*

### MscI Component List

NEB Part Number	Component Description	Lot Number	Individual QC Result
R0534LVIAL	MscI	10279763	Pass
B6004SVIAL	rCutSmart™ Buffer	10276545	Pass

### Assay Name/Specification

**Lot # 10282913**

#### Exonuclease Activity (Radioactivity Release)

A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 50 units of MscI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.

**Pass**

#### Ligation and Recutting (Terminal Integrity)

After a 20-fold over-digestion of Lambda DNA with MscI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with MscI.

**Pass**

#### Non-Specific DNase Activity (16 Hour)

A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 25 units of MscI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.

**Pass**

#### Protein Purity Assay (SDS-PAGE)

MscI is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.

**Pass**

Assay Name/Specification	Lot # 10282913
<b>qPCR DNA Contamination (E. coli Genomic)</b> A minimum of 5 units of MscI is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is $\leq 1$ E. coli genome.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Nancy Considine  
Production Scientist  
14 Mar 2025



Michael Tonello  
Packaging Quality Control Inspector  
31 Mar 2025