

New England Biolabs Certificate of Analysis

Product Name: *AhdI*
Catalog Number: *R0584L*
Concentration: *10,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot Number: *10030385*
Expiration Date: *11/2020*
Storage Temperature: *-20°C*
Storage Conditions: *50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0584S/L v2.0*

AhdI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0584LVIAL	AhdI	10030386	Pass
B7204SVIAL	CutSmart® Buffer	10021119	Pass

Assay Name/Specification	Lot # 10030385
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 30 units of AhdI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with AhdI, 95% can be recut with AhdI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of Lambda DNA and a minimum of 10 units of AhdI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
16 Oct 2018



Michael Tonello
Packaging Quality Control Inspector
07 Dec 2018