

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Ecil
Catalog Number: R0590L
Concentration: 2,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10078780
Expiration Date: 07/2021
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 μg/ml BSA

Specification Version: PS-R0590S/L v1.0

Ecil Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R0590LVIAL	Ecil	10078779	Pass	
B7204SVIAL	CutSmart® Buffer	10079734	Pass	

Assay Name/Specification	Lot # 10078780
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 2 units of Ecil incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 2-fold over-digestion of Lambda DNA with Ecil, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Ecil.	Pass
Non-Specific DNase Activity (16 hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 2 Units of Ecil incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	Pass
Protein Purity Assay (SDS-PAGE)	Pass



R0590L / Lot: 10078780

Page 1 of 2

Assay Name/Specification	Lot # 10078780
Ecil is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Penghaa Zhang Production Scientist

10 Sep 2020

Josh Hersey

Packaging Quality Control Inspector

10 Sep 2020



R0590L / Lot: 10078780

Page 2 of 2