

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Nb.BbvCl
Catalog Number: R0631L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to convert 1 µg

of supercoiled pUB DNA to open circular form in 1 hour at 37°C in a

total reaction volume of 50 μl.

Packaging Lot Number: 10089553
Expiration Date: 11/2022
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 μg/ml BSA

Specification Version: PS-R0631S/L v2.0

Nb.BbvCl Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R0631LVIAL	Nb.BbvCl	10089552	Pass	
B7204SVIAL	CutSmart® Buffer	10089402	Pass	

Assay Name/Specification	Lot # 10089553
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and	
double-stranded [ ³H] E. coli DNA and a minimum of 30 units of Nb.BbvCl incubated	
for 4 hours at 37°C releases <0.1% of the total radioactivity.	
Non-Specific DNase Activity (16 hour)	Pass
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pUB DNA and a minimum of 10	
units of Nb.BbvCl incubated for 16 hours at 37°C results in a DNA pattern free of	
detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE:	
although no nuclease degradation is detected under these conditions, extended	
incubations and/or high concentrations of this enzyme may result in star activity.	
See the product FAQ for recommended reaction conditions for this enzyme.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



R0631L / Lot: 10089553

Page 1 of 2



Penghua Zhang Production Scientist 02 Dec 2020

Michael Tonello

Packaging Quality Control Inspector

02 Dec 2020