

New England Biolabs Certificate of Analysis

Product Name: ApeKI
Catalog Number: R0643S
Concentration: 5,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in NEBuffer r3.1 in 1 hour at 75°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10184400
Expiration Date: 03/2025
Storage Temperature: -20°C
Storage Conditions: 10 mM Tris-HCl, 300 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 500 µg/ml rAlbumin (pH 7.4 @ 25°C)
Specification Version: PS-R0643S/L v2.0


ApeKI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0643SVIAL	ApeKI	10184084	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10178017	Pass
B6003SVIAL	NEBuffer™ r3.1	10182163	Pass

Assay Name/Specification	Lot # 10184400
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer™ r3.1 containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 15 units of ApeKI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Testing (15 minute Digest) A 50 µl reaction in NEBuffer™ r3.1 containing 1 µg of Lambda DNA and 1 µl of ApeKI incubated for 15 minutes at 75°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 5-fold over-digestion of Lambda DNA with ApeKI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with ApeKI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer™ r3.1 containing 1 µg of Lambda DNA and a minimum of	Pass


Assay Name/Specification	Lot # 10184400
<p>10 units of ApeKI incubated for 16 hours at 75°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	
<p>Protein Purity Assay (SDS-PAGE) ApeKI is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.</p>	Pass
<p>qPCR DNA Contamination (E. coli Genomic) A minimum of 5 units of ApeKI is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



YunJie Sun
Production Scientist
24 Mar 2023



Josh Hersey
Packaging Quality Control Inspector
07 Apr 2023