

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:BssSαICatalog Number:R0680SConcentration:10,000 U/mI

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Lot Number: 1000863:
Expiration Date: 04/2020
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl , 10 mM Tris-HCl (pH 7.4), 1 mM DTT , 0.1 mM EDTA , 50 %

Glycerol , 500 μg/ml BSA

Specification Version: PS-R0680S/L v1.0

| BssSal Component List | | | | |
|------------------------|-----------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| R0680SVIAL | BssS(alpha)I | 0061804 | Pass | |
| B7204SVIAL | CutSmart® Buffer | 3081804 | Pass | |

| Assay Name/Specification | Lot # 10008632 |
|--|----------------|
| Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and | Pass |
| double-stranded [³H] E. coli DNA and a minimum of 100 units of BssSαl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity. | |
| Functional Test (15 minute Digest) A 50 μl reaction in CutSmart [™] Buffer containing 1 μg of Lambda DNA and 1 μl of BssSαl incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis. | Pass |
| Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with BssSαI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BssSαI. | Pass |
| Non-Specific DNase Activity (16 hour) | Pass |
| A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 10 units of BssSαl incubated for 16 hours at 37°C results in a DNA pattern free of | |
| detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: | |



R0680S / Lot: 10008632

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Penghua Zhang Production Scientist

26 Jun 2018

Michael Tonello

Packaging Quality Control Inspector

26 Jun 2018



R0680S / Lot: 10008632

Page 2 of 2