

## New England Biolabs Certificate of Analysis

**Product Name:** Bccl  
**Catalog Number:** R0704L  
**Concentration:** 10,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10064537  
**Expiration Date:** 01/2022  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0704S/L v2.0

Bccl Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0704LVIAL	Bccl	10064536	Pass
B7204SVIAL	CutSmart® Buffer	10064409	Pass

Assay Name/Specification	Lot # 10064537
<p><b>Exonuclease Activity (Radioactivity Release)</b>            A 50 µl reaction in CutSmart® Buffer containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 10 units of Bccl incubated for 4 hours at 37°C releases &lt;0.3% of the total radioactivity.</p>	Pass
<p><b>Protein Purity Assay (SDS-PAGE)</b>            Bccl is &gt;95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	Pass
<p><b>Non-Specific DNase Activity (16 hour)</b>            A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 10 Units of Bccl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.</p>	Pass
<p><b>Ligation and Recutting (Terminal Integrity)</b>            After a 2-fold over-digestion of pXba DNA with Bccl, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, &gt;95%</p>	Pass

Assay Name/Specification	Lot # 10064537
can be recut with Bccl.	

This product has been tested and shown to be in compliance with all specifications.



Penghua Zhang  
Production Scientist  
14 Jan 2020



Jay Minichiello  
Packaging Quality Control Inspector  
10 Feb 2020