

PREDISCOVERY 240 County Road

BY GENUINE Ipswich, MA 01938-2723

Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Bccl
Catalog Number: R0704L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10152543
Expiration Date: 06/2024
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 μg/ml BSA

Specification Version: PS-R0704S/L v2.0

Bccl Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0704LVIAL	Bccl	10152542	Pass
B6004SVIAL	rCutSmart™ Buffer	10150374	Pass

Assay Name/Specification	Lot # 10152543
Non-Specific DNase Activity (16 hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 10 Units of Bccl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	Pass
Ligation and Recutting (Terminal Integrity) After a 2-fold over-digestion of pXba DNA with Bccl, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Bccl.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart® Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 10 units of Bccl incubated for 4 hours at 37°C releases <0.3% of the total radioactivity.	Pass
Protein Purity Assay (SDS-PAGE)	Pass



R0704L / Lot: 10152543

Page 1 of 2



Assay Name/Specification	Lot # 10152543
Bccl is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Penghaa Zhang Production Scientist

15 Jun 2022

Erin Varney

Packaging Quality Control Inspector

15 Jun 2022



R0704L / Lot: 10152543

Page 2 of 2