

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Pstl-HF®
Catalog Number: R3140T
Concentration: 100,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Lot Number: 10035461
Expiration Date: 01/2021
Storage Temperature: -20°C

Storage Conditions: 250 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol,

0.15 % Triton X-100, 200 μg/ml BSA, (pH 7.4 @ 25°C)

Specification Version: PS-R3140T/M v2.0

Pstl-HF® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R3140TVIAL	PstI-HF®	10035462	Pass	
B7204SVIAL	CutSmart® Buffer	10042965	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10038710	Pass	

Assay Name/Specification	Lot # 10035461
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in CutSmart® Buffer containing 1 μg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 200 units of Pstl-HF® incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 100-fold over-digestion of Lambda DNA with PstI-HF®, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with PstI-HF®.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of Lambda DNA and a minimum of 200 units of Pstl-HF® incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) Pstl-HF® is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass



R3140T / Lot: 10035461

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Jianying Luo Production Scientist

01 Feb 2019

Michael Tonello

Packaging Quality Control Inspector

17 May 2019

R3140T / Lot: 10035461

Page 2 of 2