

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Pvul-HF®
Catalog Number: R3150L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10119865
Expiration Date: 02/2023
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 500 μg/ml BSA

Specification Version: PS-R3150S/L v1.0

Pvul-HF® Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
R3150LVIAL	Pvul-HF®	10100683	Pass	
B7024AVIAL	Gel Loading Dye, Purple (6X)	10108733	Pass	
B6004SVIAL	rCutSmart™ Buffer	10115113	Pass	

Assay Name/Specification	Lot # 10119865
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 200 Units of Pvul-HF™ incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 200 units of Pvul-HF™ incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pXba DNA and a minimum of 200 Units of Pvul-HF™ incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 20-fold over-digestion of pXba DNA with Pvul-HF™, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments,	Pass



R3150L / Lot: 10119865

Page 1 of 2



Assay Name/Specification	Lot # 10119865
>95% can be recut with Pvul-HF™.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Penghaa Zhang Production Scientist

13 Sep 2021

Josh Hersey

Packaging Quality Control Inspector

13 Sep 2021

