240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: PvuI-HF®

Catalog #: R3150S/L

Concentration: 20,000 units/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pXba DNA in 1 hour at 37°C in a total reaction volume

of 50  $\mu l$ .

 Lot #:
 0031610

 Assay Date:
 10/2016

 Expiration Date:
 10/2018

 Storage Temp:
 -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 500 µg/ml BSA

Specification Version: PS-R3150S/L v1.0
Effective Date: 02 Oct 2013

| Assay Name/Specification (minimum release criteria)   | Lot #0031610 |
|---|--------------|
| Endonuclease Activity (Nicking) - A 50 μl reaction in CutSmart <sup>TM</sup> Buffer containing 1 μg of supercoiled PhiX174 DNA and a minimum of 200 Units of PvuI-HF <sup>TM</sup> incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.            | Pass         |
| <b>Exonuclease Activity (Radioactivity Release)</b> - A 50 μl reaction in CutSmart™ Buffer containing 1 μg of a mixture of single and double-stranded [ <sup>3</sup> H] <i>E. coli</i> DNA and a minimum of 200 units of PvuI-HF™ incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.              | Pass         |
| <b>Ligation and Recutting (Terminal Integrity)</b> - After a 20-fold over-digestion of pXba DNA with PvuI-HF <sup>TM</sup> , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with PvuI-HF <sup>TM</sup> .                                | Pass         |
| Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart <sup>TM</sup> Buffer containing 1 µg of pXba DNA and a minimum of 200 Units of PvuI-HF <sup>TM</sup> incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass         |

<sup>\*</sup> The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

Authorized by Derek Robinson 02 Oct 2013

nqa.
ISO 9001
Registered
Quality





Inspected by Jianying Luo 18 Oct 2016