

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NotI-HF®
Catalog Number: R3189S
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of pBC4 DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10062200
Expiration Date: 06/2021
Storage Temperature: -20°C

Storage Conditions: 50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 μg/ml BSA

Specification Version: PS-R3189S/L v1.0

NotI-HF® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R3189SVIAL	Notl-HF®	10046577	Pass	
B7204SVIAL	CutSmart® Buffer	10055737	Pass	
B7024SVIAL	Gel Loading Dye, Purple (6X)	10055732	Pass	

Assay Name/Specification	Lot # 10062200
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pBC4 DNA with Notl-HF™, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments,	Pass
>95% can be recut with Notl-HF™. Non-Specific DNase Activity (16 Hour)	Pass
A 50 μl reaction in CutSmart™ Buffer containing 1 μg of pBC4 DNA and a minimum of 200 Units of NotI-HF™ incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 100 Units of Notl-HF™ incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 200 units of NotI-HF™ incubated	Pass



R3189S / Lot: 10062200

Page 1 of 2

Assay Name/Specification	Lot # 10062200
for 4 hours at 37°C releases <0.1% of the total radioactivity	

This product has been tested and shown to be in compliance with all specifications.

Stephanie Cornelio Production Scientist 12 Jun 2019 Minichiello

Packaging Quality Control Inspector 27 Dec 2019

